



State of West Virginia  
 Department of Administration  
 Purchasing Division  
 2019 Washington Street East  
 Post Office Box 50130  
 Charleston, WV 25305-0130

# Request for Quotation

RFQ NUMBER  
 PSH80195

PAGE  
 1

ADDRESS CORRESPONDENCE TO ATTENTION OF  
 ROBERTA WAGNER  
 304-558-0067

RFQ COPY  
 TYPE NAME/ADDRESS HERE

PURCHASER

SHIP TO

HEALTH AND HUMAN RESOURCES  
 PINECREST HOSPITAL  
 105 SOUTH EISENHOWER DRIVE  
  
 BECKLEY, WV  
 25801 304-256-6614

DATE PRINTED	TERMS OF SALE	SHIP VIA	F.O.B.	FREIGHT TERMS
09/07/2007				
BID OPENING DATE: 09/21/2007		BID OPENING TIME 01:30PM		

LINE	QUANTITY	UOP	CAT NO	ITEM NUMBER	UNIT PRICE	AMOUNT
ADDENDUM NO. 4 1. TO CLARIFY THAT THE BID OPENING IS ON SEPTEMBER 21, 2007. 2. ADDENDUM ACKNOWLEDGEMENT IS ATTACHED. THIS DOCUMENT SHOULD BE SIGNED AND RETURNED WITH YOUR BID. FAILURE TO SIGN AND RETURN MAY RESULT IN DISQUALIFICATION OF YOUR BID. END OF ADDENDUM NO. 4						
0001	125	EA		410-03		
ELECTRIC HOME CARE BED, FOR A MATTRESS WIDTH 80"X36"						

SEE REVERSE SIDE FOR TERMS AND CONDITIONS

SIGNATURE \_\_\_\_\_ TELEPHONE \_\_\_\_\_ DATE \_\_\_\_\_

TITLE \_\_\_\_\_ FEIN \_\_\_\_\_ ADDRESS CHANGES TO BE NOTED ABOVE

WHEN RESPONDING TO RFQ, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'

## ADDENDUM #3

RFQ#PSH80195

## TO RESPOND TO VENDOR QUESTIONS AS FOLLOWS:

1. The rails specified on this RFQ will not work with the beds listed. Are the rails being ordered to be used with the beds being order?

RESPONSE: The rails being ordered are to be used with the beds listed on this RFQ. After further research, we have found that the vendor is correct. The rails being bid will not work with the beds. Therefore, please make the following changes:

**FROM: DELETE - Item #3 - 35 sets of Invacare Model RAILAF-DLX or equal  
3 position side rails and  
DELETE - Item #4 - 10 sets of Invacare Model III6808-A or equal 3/4  
side rails.**

**TO: ADD - ITEM #5 - 45 SETS OF INVACARE MODEL 6632 ASSIST  
RAILS OR EQUAL.**

- \* ALL RAILS MUST BE RUST RESISTANT.
- \* ALL RAILS MUST BE ABLE TO PIVOT  
AND/OR ADJUST TO ALLOW PATIENT  
TO ENTER OR EXIT THE BED.
- \* ALL RAILS MUST COMPLY WITH CMS  
GUIDELINES WITH REGARD TO  
ENTRAPMENT ISSUES.

**NOT OTHER CHANGES TO THIS RFQ, AT THIS TIME.**

## PURCHASING CONTINUATION SHEET

Buyer: #22	Page 3	Req or P.O. No.: psh80195
Spending Unit:		

Vendor:

Requisition No.: psh80195

## ADDENDUM ACKNOWLEDGEMENT

I hereby acknowledge receipt of the following checked addendum(s) and have made the necessary revisions to my proposal, plans and/or specifications, etc.

## Addendum No.'s:

No. 1 \_\_\_\_\_

No. 2 \_\_\_\_\_

No. 3 \_\_\_\_\_

No. 4 \_\_\_\_\_

No. 5 \_\_\_\_\_

I understand that failure to confirm the receipt of the addendum(s) may be cause for rejection of bids.

\_\_\_\_\_  
Signature\_\_\_\_\_  
Company\_\_\_\_\_  
Date