



State of West Virginia
 Department of Administration
 Purchasing Division
 2019 Washington Street East
 Post Office Box 50130
 Charleston, WV 25305-0130

Request for Quotation

RFQ NUMBER
LBS90133

PAGE
1

ADDRESS CORRESPONDENCE TO ATTENTION OF
ROBERTA WAGNER
304-558-0067

VENDOR

RFQ COPY
 TYPE NAME/ADDRESS HERE

SHIP TO

HEALTH AND HUMAN RESOURCES
 ENVIRONMENTAL CHEMISTRY LAB
 4710 CHIMNEY DRIVE
 CHARLESTON, WV
 25302 304-558-3530

DATE PRINTED	TERMS OF SALE	SHIP VIA	F.O.B	FREIGHT TERMS
05/19/2009				

BID OPENING DATE: **06/18/2009** BID OPENING TIME **01:30PM**

LINE	QUANTITY	UOP	CAT NO.	ITEM NUMBER	UNIT PRICE	AMOUNT
0001	1	EA		490-55		
GAS CHROMATOGRAPH/MASS SPECTROMETER (GC-MS)						
TO ANALYZE FOR TRIHALOMETHANES (THMS) AND VOLATILE ORGANIC CHEMICALS (VOCs) PER THE ATTACHED SPECS.						
0002	1	EA		490-55		
GAS CHROMATOGRAPH/ELECTRON CAPTURE DETECTOR (GC-ECD)						
TO ANALYZE FOR HALOACETIC ACIDS (HAAS) IN DRINKING WATER PER THE ATTACHED SPECS.						
0003	1	EA		490-55		
ON-SITE BASIC USER TRAINING FOR GC-MS SPECTROMETER						
(1 DAY AT INSTALLATION OF EQUIPMENT.)						
0004	3	EA		490-55		
ON SITE BASIC USER TRAINING FOR GC-ECD DETECTOR						
(3 DAYS AT INSTALLATION OF EQUIPMENT.)						

SEE REVERSE SIDE FOR TERMS AND CONDITIONS

SIGNATURE	TELEPHONE	DATE
TITLE	FEIN	ADDRESS CHANGES TO BE NOTED ABOVE

WHEN RESPONDING TO RFQ, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'

GENERAL TERMS & CONDITIONS
REQUEST FOR QUOTATION (RFQ) AND REQUEST FOR PROPOSAL (RFP)

1. Awards will be made in the best interest of the State of West Virginia.
2. The State may accept or reject in part, or in whole, any bid.
3. All quotations are governed by the *West Virginia Code* and the *Legislative Rules* of the Purchasing Division.
4. Prior to any award, the apparent successful vendor must be properly registered with the Purchasing Division and have paid the required \$125 fee.
5. All services performed or goods delivered under State Purchase Order/Contracts are to be continued for the term of the Purchase Order/Contracts, contingent upon funds being appropriated by the Legislature or otherwise being made available. In the event funds are not appropriated or otherwise available for these services or goods, this Purchase Order/Contract becomes void and of no effect after June 30.
6. Payment may only be made after the delivery and acceptance of goods or services.
7. Interest may be paid for late payment in accordance with the *West Virginia Code*.
8. Vendor preference will be granted upon written request in accordance with the *West Virginia Code*.
9. The State of West Virginia is exempt from federal and state taxes and will not pay or reimburse such taxes.
10. The Director of Purchasing may cancel any Purchase Order/Contract upon 30 days written notice to the seller.
11. The laws of the State of West Virginia and the *Legislative Rules* of the Purchasing Division shall govern all rights and duties under the Contract, including without limitation the validity of this Purchase Order/Contract.
12. Any reference to automatic renewal is hereby deleted. The Contract may be renewed only upon mutual written agreement of the parties.
13. **BANKRUPTCY:** In the event the vendor/contractor files for bankruptcy protection, this Contract may be deemed null and void, and terminated without further order.
14. **HIPAA BUSINESS ASSOCIATE ADDENDUM:** The West Virginia State Government HIPAA Business Associate Addendum (BAA), approved by the Attorney General, and available online at the Purchasing Division's web site (<http://www.state.wv.us/admin/purchase/vrc/hipaa.htm>) is hereby made part of the agreement. Provided that, the Agency meets the definition of a Cover Entity (45 CFR §160.103) and will be disclosing Protected Health Information (45 CFR §160.103) to the vendor.
15. **WEST VIRGINIA ALCOHOL & DRUG-FREE WORKPLACE ACT:** If this Contract constitutes a public improvement construction contract as set forth in Article 1D, Chapter 21 of the West Virginia Code ("The West Virginia Alcohol and Drug-Free Workplace Act"), then the following language shall hereby become part of this Contract: "The contractor and its subcontractors shall implement and maintain a written drug-free workplace policy in compliance with the West Virginia Alcohol and Drug-Free Workplace Act, as set forth in Article 1D, Chapter 21 of the West Virginia Code. The contractor and its subcontractors shall provide a sworn statement in writing, under the penalties of perjury, that they maintain a valid drug-free work place policy in compliance with the West Virginia and Drug-Free Workplace Act. It is understood and agreed that this Contract shall be cancelled by the awarding authority if the Contractor: 1) Fails to implement its drug-free workplace policy; 2) Fails to provide information regarding implementation of the contractor's drug-free workplace policy at the request of the public authority; or 3) Provides to the public authority false information regarding the contractor's drug-free workplace policy."

INSTRUCTIONS TO BIDDERS

1. Use the quotation forms provided by the Purchasing Division.
2. **SPECIFICATIONS:** Items offered must be in compliance with the specifications. Any deviation from the specifications must be clearly indicated by the bidder. Alternates offered by the bidder as **EQUAL** to the specifications must be clearly defined. A bidder offering an alternate should attach complete specifications and literature to the bid. The Purchasing Division may waive minor deviations to specifications.
3. Complete all sections of the quotation form.
4. Unit prices shall prevail in case of discrepancy.
5. All quotations are considered F.O.B. destination unless alternate shipping terms are clearly identified in the quotation.
6. **BID SUBMISSION:** All quotations must be delivered by the bidder to the office listed below prior to the date and time of the bid opening. Failure of the bidder to deliver the quotations on time will result in bid disqualifications: Department of Administration, Purchasing Division, 2019 Washington Street East, P.O. Box 50130, Charleston, WV 25305-0130



State of West Virginia
 Department of Administration
 Purchasing Division
 2019 Washington Street East
 Post Office Box 50130
 Charleston, WV 25305-0130

Request for Quotation

RFQ NUMBER
LBS90133

PAGE
2

ADDRESS CORRESPONDENCE TO ATTENTION OF
ROBERTA WAGNER
304-558-0067

RFQ COPY

TYPE NAME/ADDRESS HERE

SHIP TO

HEALTH AND HUMAN RESOURCES
 ENVIRONMENTAL CHEMISTRY LAB
 4710 CHIMNEY DRIVE
 CHARLESTON, WV
 25302 304-558-3530

DATE PRINTED	TERMS OF SALE	SHIP VIA	F.O.B.	FREIGHT TERMS
05/19/2009				

BID OPENING DATE: **06/18/2009** BID OPENING TIME **01:30PM**

LINE	QUANTITY	UOP	CAT NO.	ITEM NUMBER	UNIT PRICE	AMOUNT
0005	4	EA		490-55		
<p>ADVANCED TRAINING COURSE FOR THE GC-MS SPECTROMETER</p> <p>(4 DAYS)</p> <p>THE AWARD WILL BE MADE TO THE VENDOR WITH THE LOWEST OVERALL TOTAL COST OF THE EQUIPMENT WHICH MEETS ALL REQUESTED SPECIFICATIONS AND REQUIREMENTS. PAYMENT WILL BE MADE IN ARREARS.</p> <p>CANCELLATION: THE DIRECTOR OF PURCHASING RESERVES THE RIGHT TO CANCEL THIS CONTRACT IMMEDIATELY UPON WRITTEN NOTICE TO THE VENDOR IF THE COMMODITIES AND/OR SERVICES SUPPLIED ARE OF AN INFERIOR QUALITY OR DO NOT CONFORM TO THE SPECIFICATIONS OF THE BID AND CONTRACT HEREIN.</p> <p>BANKRUPTCY: IN THE EVENT THE VENDOR/CONTRACTOR FILES FOR BANKRUPTCY PROTECTION, THIS CONTRACT IS AUTOMATICALLY NULL AND VOID, AND IS TERMINATED WITHOUT FURTHER ORDER.</p> <p>INQUIRIES: WRITTEN QUESTIONS SHALL BE ACCEPTED THROUGH CLOSE OF BUSINESS ON 6/2/2009. QUESTIONS MAY BE SENT VIA USPS, FAX, COURIER OR E-MAIL. IN ORDER TO ASSURE NO VENDOR RECEIVES AN UNFAIR ADVANTAGE, NO SUBSTANTIVE QUESTIONS WILL BE ANSWERED ORALLY. IF POSSIBLE, E-MAIL QUESTIONS ARE PREFERRED. ADDRESS INQUIRIES TO:</p> <p>ROBERTA WAGNER DEPARTMENT OF ADMINISTRATION</p>						

SEE REVERSE SIDE FOR TERMS AND CONDITIONS

SIGNATURE	TELEPHONE	DATE
TITLE	FEIN	ADDRESS CHANGES TO BE NOTED ABOVE

WHEN RESPONDING TO RFQ, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'



State of West Virginia
 Department of Administration
 Purchasing Division
 2019 Washington Street East
 Post Office Box 50130
 Charleston, WV 25305-0130

Request for Quotation

RFQ NUMBER
LBS90133

PAGE
3

ADDRESS CORRESPONDENCE TO ATTENTION OF
**ROBERTA WAGNER
 304-558-0067**

VENDOR

RFQ COPY
 TYPE NAME/ADDRESS HERE

SHIP TO

HEALTH AND HUMAN RESOURCES
 ENVIRONMENTAL CHEMISTRY LAB
 4710 CHIMNEY DRIVE
 CHARLESTON, WV
 25302 304-558-3530

DATE PRINTED	TERMS OF SALE	SHIP VIA	F.O.B.	FREIGHT TERMS
05/19/2009				

BID OPENING DATE: **06/18/2009** BID OPENING TIME **01:30PM**

LINE	QUANTITY	UOP	CAT. NO.	ITEM NUMBER	UNIT PRICE	AMOUNT
PURCHASING DIVISION 2019 WASHINGTON STREET, EAST CHARLESTON, WV 25311 FAX: 304-558-4115 E-MAIL: ROBERTA.A.WAGNER@WV.GOV						
THE MODEL/BRAND/SPECIFICATIONS NAMED HEREIN ESTABLISH THE ACCEPTABLE LEVEL OF QUALITY ONLY AND ARE NOT INTENDED TO REFLECT A PREFERENCE OR FAVOR ANY PARTICULAR BRAND OR VENDOR. VENDORS WHO ARE BIDDING ALTERNATES SHOULD SO STATE AND INCLUDE PERTINENT LITERATURE AND SPECIFICATIONS. FAILURE TO PROVIDE INFORMATION FOR ANY ALTERNATES MAY BE GROUNDS FOR REJECTION OF THE BID. THE STATE RESERVES THE RIGHT TO WAIVE MINOR IRREGULARITIES IN BIDS OR SPECIFICATIONS IN ACCORDANCE WITH SECTION 148-1-4(F) OF THE WEST VIRGINIA LEGISLATIVE RULES AND REGULATIONS.						
NOTICE						
A SIGNED BID MUST BE SUBMITTED TO:						
DEPARTMENT OF ADMINISTRATION PURCHASING DIVISION BUILDING 15 2019 WASHINGTON STREET, EAST CHARLESTON, WV 25305-0130						
PLEASE NOTE: A CONVENIENCE COPY WOULD BE APPRECIATED.						
THE BID SHOULD CONTAIN THIS INFORMATION ON THE FACE OF						

SEE REVERSE SIDE FOR TERMS AND CONDITIONS

SIGNATURE	TELEPHONE	DATE
TITLE	FEIN	ADDRESS CHANGES TO BE NOTED ABOVE

WHEN RESPONDING TO RFQ, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'



State of West Virginia
 Department of Administration
 Purchasing Division
 2019 Washington Street East
 Post Office Box 50130
 Charleston, WV 25305-0130

Request for Quotation

RFQ NUMBER
LBS90133

PAGE
4

ADDRESS CORRESPONDENCE TO ATTENTION OF
ROBERTA WAGNER
304-558-0067

VENDOR

RFQ COPY
 TYPE NAME/ADDRESS HERE

SHIP TO

HEALTH AND HUMAN RESOURCES
 ENVIRONMENTAL CHEMISTRY LAB
 4710 CHIMNEY DRIVE
 CHARLESTON, WV
 25302 304-558-3530

DATE PRINTED	TERMS OF SALE	SHIP VIA	F.O.B.	FREIGHT TERMS
05/19/2009				

BID OPENING DATE: **06/18/2009** BID OPENING TIME **01:30PM**

LINE	QUANTITY	UOP	CAT NO.	ITEM NUMBER	UNIT PRICE	AMOUNT
THE ENVELOPE OR THE BID MAY NOT BE CONSIDERED: SEALED BID BUYER:-----RW/FILE 22----- RFQ. NO.:-----LBS90133----- BID OPENING DATE:-----6/18/2009----- BID OPENING TIME:-----1:30 PM----- PLEASE PROVIDE A FAX NUMBER IN CASE IT IS NECESSARY TO CONTACT YOU REGARDING YOUR BID: ----- CONTACT PERSON (PLEASE PRINT CLEARLY): ----- ***** THIS IS THE END OF RFQ LBS90133 ***** TOTAL: _____						

SEE REVERSE SIDE FOR TERMS AND CONDITIONS

SIGNATURE	TELEPHONE	DATE
TITLE	FEIN	ADDRESS CHANGES TO BE NOTED ABOVE

WHEN RESPONDING TO RFQ, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'

Gas Chromatograph – Mass Spectrometer

SPECIFICATIONS AND REQUIREMENTS:

1. The Environmental Chemistry Laboratory section of the Office of Laboratory Services is requesting to purchase a Gas Chromatograph – Mass Spectrometer (GC-MS) instrument to analyze for Trihalomethanes (THMs) and Volatile Organic Chemicals (VOCs) in drinking water to fulfill requirements of United States Environmental Protection Agency (EPA) regulations for compliance monitoring of public water systems. Compliance monitoring testing is a requirement of a state's Principal State Laboratory mandated under federal code at 40 CFR 142.10(b)(4) for the state to maintain Primacy over its Drinking Water Program. The instrument will also serve the purpose of protecting public health by providing testing capabilities to private well owners. This instrument is to be installed for use by the Environmental Chemistry Laboratory located at 4710 Chimney Drive, Suite G, Charleston, WV 25302.
2. The technology of this GC-MS instrument must be compatible with and fulfill the requirements of United States Environmental Protection Agency (EPA) method 524.2 (1995, Revision 4.1) for THMs and VOCs, and the resolution and sensitivity capability must be able to meet or exceed the limits of detection and minimum reporting levels required for these regulated compounds as stipulated under the federal Safe Drinking Water Act, and the Stage 2 Disinfection Byproduct Rule.
3. The GC-MS Instrument must be fully automated for analysis with a system controller that is loaded with the necessary software which is compatible with Windows XP. System controller software must be able to export data to the existing Laboratory Information Management System (LIMS).
4. The GC-MS Instrument system controller software must be able to fulfill all of the analytical and quality control requirements stated in the EPA Method 524.2 (Rev. 4.1) as well as the pertinent requirements listed in Chapter IV of the EPA "Manual for the Certification of Laboratories Analyzing Drinking Water, Fifth edition, January 2005", for the analysis of THMs and VOCs in drinking water.
5. GC-MS Instrument must be fully compatible with a system software controlled autosampler that is integral with a Purge and Trap concentrator sampling system manufactured by OI Analytical. The autosampler is Model 4551-A. The Purge and Trap is the Eclipse Model 4660. The OI Analytical autosampler and purge and trap have already been purchased separately.
6. The GC-MS Instrument must come with a temperature-programmable split/splitless capillary column injector compatible with a Purge and Trap sampling interface.
7. The GC-MS Instrument must come with a column for method 524.2 volatiles analysis.
8. The GC-MS Instrument must come equipped with a quadrupole mass spectrometer.
9. The GC-MS Instrument must come equipped with a pre-quad filter, or fully equivalent technology, to protect the quadrupole from contamination.

10. The GC-MS Instrument must be capable of mass scan rates of at least 12,500 amu/sec and 65 scans/sec.
11. The GC-MS Instrument must be capable of detecting a wide mass range of at least 1.5-1050 u (Daltons).
12. The GC-MS Instrument must come equipped with a 179 L/sec turbomolecular pump or better.
13. The GC-MS Instrument must be capable of being attached to a nitrogen purge gas stream to prohibit oxygen from entering the system during venting, or provide equivalent technology, to minimize contamination in case of a power outage.
14. The GC-MS Instrument oven must be capable of a rapid cool-down rate, from 450°C to 50°C, in less than 2 minutes to facilitate fast analytical cycles.
15. The GC-MS Instrument must come equipped with a programmable pneumatic control system that controls and monitors all injector, detector, and auxiliary gases electronically, or fully equivalent technology.
16. The GC-MS Instrument system must operate on 110-120 V AC.

Installation Requirements:

1. Vendor must install the GC-MS Instrument system in the Environmental Chemistry Laboratory in the Elk Office Center building at 4710 Chimney Drive, Suite G, Charleston, WV 25302 .
2. Vendor must provide to the Environmental Chemistry Laboratory (ECL) Program Manager all relevant information concerning the installation in a documented form at least 2 weeks prior to the scheduled installation.
3. At the time of the GC-MS Instrument installation, Vendor must provide to the ECL Program Manager: All relevant system manuals for hardware components; system and application software documentation; a parts, supplies, accessories catalog.
4. Vendor must include all necessary cables and fittings and other costs for installation in the submitted bid price.
5. Installation and on-site training must be completed within 90 days of delivery date.

Training Requirements:

1. Vendor must provide, upon completion of installation, basic on-site training for Environmental Chemistry Laboratory personnel on the operation and user maintenance requirements of the GC-MS Instrument system.
2. On-site training provided by Vendor must last for at least 1 day, following GC-MS Instrument system installation.
3. All costs incurred by the Vendor including travel, lodging, and living expenses necessary to provide this basic training shall be included in the bid price.
4. Vendor must also provide an advanced and complete training course that includes the theory of GC-MS, workshops demonstrating all aspects of the proprietary software

capabilities, and hands-on lab work. This course must last at least 4 days. Vendor must include the cost of this course in the bid price.

Warranty Requirements:

1. Vendor must include in the total price of the equipment at least 1-year factory warranty covering all system components.
2. Software support must be included as part of one-year warranty. Vendor must agree to provide software support for any subsequent service agreements that may follow the original one-year standard warranty.
3. Warranty must include on-site service including labor, travel time, and expenses with a 24-hour on-site response time at no extra cost to maintain the specifications listed in this bid and the Vendor's product specifications.
4. Vendor must include all warranty information including terms and conditions upon delivery.
5. Warranty must begin upon completion of installation.
6. Service provided under warranty must be provided directly by an employee of the Vendor's company exclusive of contracted third parties.
7. Vendor must agree to offer the first one year extended service warranty after the initial 1 year warranty expires at a cost indicated on the bid response page.
8. For a period of ten years commencing with the end of the one year warranty period following installation, the Vendor must make available a one-year service period for preventive maintenance and emergency service on both system hardware and software on a reoccurring annual basis.

Delivery Requirements:

1. The GC-MS Instrument and its components must be shipped for "inside delivery" by freight delivery company.
2. The GC-MS Instrument and its components must be shipped "F.O.B. Destination" unless otherwise stated in quote by Vendor. Any shipping and handling requirements must be stated in Vendor's quote.
3. Delivery must be within 90 days after receipt of order.

Conformance Requirements:

1. The GC-MS Instrument system shall be operated for a period of 30 days following installation and must be found to conform to the herein listed specifications.
2. Vendor must agree to accept the return of the GC-MS Instrument system should the GC-MS Instrument system fail to conform to any required specification within this 30 day period.

3. Should the GC-MS Instrument system be returned to the Vendor, the Vendor must agree to provide a full refund of the bid price of the GC-MS Instrument system to the State of West Virginia.

Gas Chromatograph – Electron Capture Detector

SPECIFICATIONS AND REQUIREMENTS:

1. The Environmental Chemistry Laboratory section of the Office of Laboratory Services is requesting for purchase a Gas Chromatograph – Electron Capture Detector (GC-ECD) instrument to analyze for Haloacetic Acids (HAA5) in drinking water to fulfill requirements of United States Environmental Protection Agency (EPA) regulations for compliance monitoring of public water systems. Compliance monitoring testing is a requirement of a state's Principal State Laboratory mandated under federal code at 40 CFR 142.10(b)(4) for the state to maintain Primacy over its Drinking Water Program. The instrument will also serve the purpose of protecting public health by providing testing capabilities to private well owners. This instrument is to be installed for use by the Environmental Chemistry Laboratory at 4710 Chimney Drive, Suite G, Charleston, WV 25302.
2. The technology of this GC-ECD instrument must be compatible with and fulfill the requirements of United States Environmental Protection Agency (EPA) method 552.3 (2003, Revision 1.0) for Haloacetic Acids (HAA5) and the resolution and sensitivity capability must be able to meet or exceed the limits of detection and minimum reporting levels required for these regulated compounds as stipulated under the federal Safe Drinking Water Act, and the Stage 2 Disinfection Byproduct Rule.
3. The GC-ECD Instrument must be fully automated for analysis with a system controller that is loaded with the necessary software which is compatible with Windows XP. System controller software must be able to export data to the existing Laboratory Information Management System (LIMS).
4. The GC-ECD Instrument system controller software must be able to fulfill all of the analytical and quality control requirements stated in the EPA Method 552.3 (Rev. 1.0) as well as the pertinent requirements listed in Chapter IV of the EPA "Manual for the Certification of Laboratories Analyzing Drinking Water, Fifth edition, January 2005", for the analysis of HAA5 in drinking water.
5. Instrument must be equipped with a system software controlled autosampler.
6. The GC-ECD Instrument must come with a temperature-programmable split / splitless capillary column injector.
7. The GC-ECD Instrument must come equipped with a Primary GC Column for method 552.3 volatiles analysis. This is a DB-1701, 30-meter length, 0.25-mm i.d., 0.25- μ m film, fused silica capillary with chemically bonded (14% cyanopropylphenyl-methylpolysiloxane), or equivalent bonded, fused silica column.

8. The GC-ECD Instrument must come equipped with a Secondary Confirmation GC Column. This is a DB-5.625, 30-meter length, 0.25-mm i.d., 0.25- μ m film, fused silica capillary with chemically bonded ("equivalent to" 5% phenylmethylpolysiloxane), or equivalent bonded, fused silica column.
9. The GC-ECD Instrument must come equipped with a low volume, micro, Electron Capture Detector for each of the Primary and Secondary Confirmation columns.
10. The GC-ECD Instrument oven must be capable of a rapid cool-down rate from 450 °C to 50 °C in less than 2 minutes to facilitate fast analytical cycle times.
11. The GC-ECD Instrument must come equipped with a programmable pneumatic control system, or fully equivalent technology, that controls and monitors all injector, detector, and auxiliary gases electronically.
12. The GC-ECD Instrument system must operate on 110-120 V AC.

Installation Requirements:

1. Vendor must install the GC-ECD Instrument system in the Environmental Chemistry Laboratory in the Elk Office Center building at 4710 Chimney Drive, Suite G, Charleston, WV 25302 .
2. Vendor must provide to the Environmental Chemistry Laboratory (ECL) Program Manager all relevant information concerning the installation in a documented form at least 2 weeks prior to the scheduled installation.
3. At the time of the GC-ECD Instrument installation, Vendor must provide to the ECL Program Manager: All relevant system manuals for hardware components; system and application software documentation; a parts, supplies, accessories catalog.
4. Vendor must include all necessary cables and fittings and other costs for installation in the submitted bid price.
5. Installation and on-site training must be completed within 90 days of delivery date.

Training Requirements:

1. Vendor must provide on-site training for Environmental Chemistry Laboratory personnel on the operation and user maintenance requirements of the GC-ECD, following completion of installation.
2. Vendor must provide on-site training relevant to the acceptable performance of the EPA method 552.3.
3. On-site training provided by Vendor must be for a minimum of 3 days following Instrument system installation.
4. All costs incurred by the Vendor including travel, lodging, and living expenses necessary to provide this training shall be included in the bid price.

Warranty Requirements:

1. Vendor must include in the total price of the equipment at least 1-year factory warranty covering all system components.

2. Software support must be included as part of one-year warranty. Vendor must agree to provide software support for any subsequent service agreements that may follow the original one-year standard warranty.
3. Warranty must include on-site service including labor, travel time, and expenses with a 24-hour on-site response time at no extra cost to maintain the specifications listed in this bid and the Vendor's product specifications.
4. Vendor must include all warranty information including terms and conditions upon delivery.
5. Warranty to begin upon completion of installation.
6. Service provided under warranty must be provided directly by an employee of the Vendor's company exclusive of contracted third parties.
7. Vendor must agree to offer the first one year extended service warranty after the initial 1 year warranty expires at a cost indicated on the bid response page.
8. For a period of ten years commencing with the end of the one year warranty period following installation, the Vendor must make available a one-year service period for preventive maintenance and emergency service on both system hardware and software on a reoccurring annual basis.

Delivery Requirements:

1. The GC-ECD Instrument and its components must be shipped for "inside delivery" by freight delivery company.
2. The GC-ECD Instrument and its components are to be shipped "F.O.B. Destination" unless otherwise stated in quote by Vendor. Any shipping and handling requirements must be stated in Vendor's quote.
3. Delivery must be within 90 days after receipt of order.

Conformance Requirements:

1. The GC-ECD system must be operated for a period of 30 days following installation and must be found to conform to the herein listed specifications.
2. Vendor must agree to accept the return of the GC-ECD system should the GC-ECD system fail to conform to any required specification within this 30 day period.
3. Should the GC-ECD system be returned to the Vendor, the Vendor must agree to provide a full refund of the bid price of the GC-ECD system to the State of West Virginia.

RFQ COST SHEET

Bidders shall provide a cost for the following:

Gas Chromatograph – Mass Spectrometer (GC-MS) \$ _____

On-site Basic User Training (1 day at installation of equipment) \$ _____

Advanced Training Course (4 days) \$ _____

Gas Chromatograph – Electron Capture Detector (GC-ECD) \$ _____

On-site Basic User Training (3 days at installation of equipment) \$ _____

Total Cost \$ _____

The award will be made to the vendor with the lowest overall total cost of the equipment which meets all requested specifications and requirements. Payment will be made in arrears.

Vendor Signature

Date

AGREEMENT ADDENDUM

WV-96
Rev. 10/07

In the event of conflict between this addendum and the agreement, this addendum shall control:

1. **DISPUTES** - Any references in the agreement to arbitration or to the jurisdiction of any court are hereby deleted. Disputes arising out of the agreement shall be presented to the West Virginia Court of Claims.
2. **HOLD HARMLESS** - Any clause requiring the Agency to indemnify or hold harmless any party is hereby deleted in its entirety.
3. **GOVERNING LAW** - The agreement shall be governed by the laws of the State of West Virginia. This provision replaces any references to any other State's governing law.
4. **TAXES** - Provisions in the agreement requiring the Agency to pay taxes are deleted. As a State entity, the Agency is exempt from Federal, State, and local taxes and will not pay taxes for any Vendor including individuals, nor will the Agency file any tax returns or reports on behalf of Vendor or any other party.
5. **PAYMENT** - Any references to prepayment are deleted. Payment will be in arrears.
6. **INTEREST** - Should the agreement include a provision for interest on late payments, the Agency agrees to pay the maximum legal rate under West Virginia law. All other references to interest or late charges are deleted.
7. **RECOUPMENT** - Any language in the agreement waiving the Agency's right to set-off, counterclaim, recoupment, or other defense is hereby deleted.
8. **FISCAL YEAR FUNDING** - Service performed under the agreement may be continued in succeeding fiscal years for the term of the agreement, contingent upon funds being appropriated by the Legislature or otherwise being available for this service. In the event funds are not appropriated or otherwise available for this service, the agreement shall terminate without penalty on June 30. After that date, the agreement becomes of no effect and is null and void. However, the Agency agrees to use its best efforts to have the amounts contemplated under the agreement included in its budget. Non-appropriation or non-funding shall not be considered an event of default.
9. **STATUTE OF LIMITATION** - Any clauses limiting the time in which the Agency may bring suit against the Vendor, lessor, individual, or any other party are deleted.
10. **SIMILAR SERVICES** - Any provisions limiting the Agency's right to obtain similar services or equipment in the event of default or non-funding during the term of the agreement are hereby deleted.
11. **ATTORNEY FEES** - The Agency recognizes an obligation to pay attorney's fees or costs only when assessed by a court of competent jurisdiction. Any other provision is invalid and considered null and void.
12. **ASSIGNMENT** - Notwithstanding any clause to the contrary, the Agency reserves the right to assign the agreement to another State of West Virginia agency, board or commission upon thirty (30) days written notice to the Vendor and Vendor shall obtain the written consent of Agency prior to assigning the agreement.
13. **LIMITATION OF LIABILITY** - The Agency, as a State entity, cannot agree to assume the potential liability of a Vendor. Accordingly, any provision limiting the Vendor's liability for direct damages to a certain dollar amount or to the amount of the agreement is hereby deleted. Limitations on special, incidental or consequential damages are acceptable. In addition, any limitation is null and void to the extent that it precludes any action for injury to persons or for damages to personal property.
14. **RIGHT TO TERMINATE** - Agency shall have the right to terminate the agreement upon thirty (30) days written notice to Vendor. Agency agrees to pay Vendor for services rendered or goods received prior to the effective date of termination.
15. **TERMINATION CHARGES** - Any provision requiring the Agency to pay a fixed amount or liquidated damages upon termination of the agreement is hereby deleted. The Agency may only agree to reimburse a Vendor for actual costs incurred or losses sustained during the current fiscal year due to wrongful termination by the Agency prior to the end of any current agreement term.
16. **RENEWAL** - Any reference to automatic renewal is hereby deleted. The agreement may be renewed only upon mutual written agreement of the parties.
17. **INSURANCE** - Any provision requiring the Agency to insure equipment or property of any kind and name the Vendor as beneficiary or as an additional insured is hereby deleted.
18. **RIGHT TO NOTICE** - Any provision for repossession of equipment without notice is hereby deleted. However, the Agency does recognize a right of repossession with notice.
19. **ACCELERATION** - Any reference to acceleration of payments in the event of default or non-funding is hereby deleted.
20. **CONFIDENTIALITY** - Any provision regarding confidentiality of the terms and conditions of the agreement is hereby deleted. State contracts are public records under the West Virginia Freedom of Information Act.
21. **AMENDMENTS** - All amendments, modifications, alterations or changes to the agreement shall be in writing and signed by both parties. No amendment, modification, alteration or change may be made to this addendum without the express written approval of the Purchasing Division and the Attorney General.

ACCEPTED BY:

STATE OF WEST VIRGINIA

VENDOR

Spending Unit: _____

Company Name: _____

Signed: _____

Signed: _____

Title: _____

Title: _____

Date: _____

Date: _____

ATTACHMENT
P.O.# LBS90133

This agreement constitutes the entire agreement between the parties, and there are no other terms and conditions applicable to the licenses granted hereunder.

Agreed

Signature Date

Title

Company Name

Signature Date

Title

Agency/Division

STATE OF WEST VIRGINIA
Purchasing Division

PURCHASING AFFIDAVIT

VENDOR OWING A DEBT TO THE STATE:

West Virginia Code §5A-3-10a provides that: No contract or renewal of any contract may be awarded by the state or any of its political subdivisions to any vendor or prospective vendor when the vendor or prospective vendor or a related party to the vendor or prospective vendor is a debtor and the debt owed is an amount greater than one thousand dollars in the aggregate.

PUBLIC IMPROVEMENT CONTRACTS & DRUG-FREE WORKPLACE ACT:

If this is a solicitation for a public improvement construction contract, the vendor, by its signature below, affirms that it has a written plan for a drug-free workplace policy in compliance with Article 1D, Chapter 21 of the **West Virginia Code**. The vendor **must** make said affirmation with its bid submission. Further, public improvement construction contract may not be awarded to a vendor who does not have a written plan for a drug-free workplace policy in compliance with Article 1D, Chapter 21 of the **West Virginia Code** and who has not submitted that plan to the appropriate contracting authority in timely fashion. For a vendor who is a subcontractor, compliance with Section 5, Article 1D, Chapter 21 of the **West Virginia Code** may take place before their work on the public improvement is begun.

ANTITRUST:

In submitting a bid to any agency for the state of West Virginia, the bidder offers and agrees that if the bid is accepted the bidder will convey, sell, assign or transfer to the state of West Virginia all rights, title and interest in and to all causes of action it may now or hereafter acquire under the antitrust laws of the United States and the state of West Virginia for price fixing and/or unreasonable restraints of trade relating to the particular commodities or services purchased or acquired by the state of West Virginia. Such assignment shall be made and become effective at the time the purchasing agency tenders the initial payment to the bidder.

I certify that this bid is made without prior understanding, agreement, or connection with any corporation, firm, limited liability company, partnership or person or entity submitting a bid for the same materials, supplies, equipment or services and is in all respects fair and without collusion or fraud. I further certify that I am authorized to sign the certification on behalf of the bidder or this bid.

LICENSING:

Vendors must be licensed and in good standing in accordance with any and all state and local laws and requirements by any state or local agency of West Virginia, including, but not limited to, the West Virginia Secretary of State's Office, the West Virginia Tax Department, West Virginia Insurance Commission, or any other state agencies or political subdivision. Furthermore, the vendor must provide all necessary releases to obtain information to enable the Director or spending unit to verify that the vendor is licensed and in good standing with the above entities.

CONFIDENTIALITY:

The vendor agrees that he or she will not disclose to anyone, directly or indirectly, any such personally identifiable information or other confidential information gained from the agency, unless the individual who is the subject of the information consents to the disclosure in writing or the disclosure is made pursuant to the agency's policies, procedures and rules. Vendor further agrees to comply with the Confidentiality Policies and Information Security Accountability Requirements, set forth in <http://www.state.wv.us/admin/purchase/privacy/noticeConfidentiality.pdf>.

Under penalty of law for false swearing (**West Virginia Code §61-5-3**), it is hereby certified that the vendor affirms and acknowledges the information in this affidavit and is in compliance with the requirements as stated.

Vendor's Name: _____
Authorized Signature: _____ Date: _____

State of West Virginia

VENDOR PREFERENCE CERTIFICATE

Certification and application* is hereby made for Preference in accordance with West Virginia Code, §5A-3-37. (Does not apply to construction contracts). West Virginia Code, §5A-3-37, provides an opportunity for qualifying vendors to request (at the time of bid) preference for their residency status. Such preference is an evaluation method only and will be applied only to the cost bid in accordance with the West Virginia Code. This certificate for application is to be used to request such preference. The Purchasing Division will make the determination of the Resident Vendor Preference, if applicable.

- 1. Application is made for 2.5% resident vendor preference for the reason checked: Bidder is an individual resident vendor and has resided continuously in West Virginia for four (4) years immediately preceding the date of this certification; or, Bidder is a partnership, association or corporation resident vendor and has maintained its headquarters or principal place of business continuously in West Virginia for four (4) years immediately preceding the date of this certification; or 80% of the ownership interest of Bidder is held by another individual, partnership, association or corporation resident vendor who has maintained its headquarters or principal place of business continuously in West Virginia for four (4) years immediately preceding the date of this certification; or, Bidder is a nonresident vendor which has an affiliate or subsidiary which employs a minimum of one hundred state residents and which has maintained its headquarters or principal place of business within West Virginia continuously for the four (4) years immediately preceding the date of this certification; or,
2. Application is made for 2.5% resident vendor preference for the reason checked: Bidder is a resident vendor who certifies that, during the life of the contract, on average at least 75% of the employees working on the project being bid are residents of West Virginia who have resided in the state continuously for the two years immediately preceding submission of this bid; or,
3. Application is made for 2.5% resident vendor preference for the reason checked: Bidder is a nonresident vendor employing a minimum of one hundred state residents or is a nonresident vendor with an affiliate or subsidiary which maintains its headquarters or principal place of business within West Virginia employing a minimum of one hundred state residents who certifies that, during the life of the contract, on average at least 75% of the employees or Bidder's affiliate's or subsidiary's employees are residents of West Virginia who have resided in the state continuously for the two years immediately preceding submission of this bid; or,
4. Application is made for 5% resident vendor preference for the reason checked: Bidder meets either the requirement of both subdivisions (1) and (2) or subdivision (1) and (3) as stated above; or,
5. Application is made for 3.5% resident vendor preference who is a veteran for the reason checked: Bidder is an individual resident vendor who is a veteran of the United States armed forces, the reserves or the National Guard and has resided in West Virginia continuously for the four years immediately preceding the date on which the bid is submitted; or,
6. Application is made for 3.5% resident vendor preference who is a veteran for the reason checked: Bidder is a resident vendor who is a veteran of the United States armed forces, the reserves or the National Guard, if, for purposes of producing or distributing the commodities or completing the project which is the subject of the vendor's bid and continuously over the entire term of the project, on average at least seventy-five percent of the vendor's employees are residents of West Virginia who have resided in the state continuously for the two immediately preceding years.

Bidder understands if the Secretary of Revenue determines that a Bidder receiving preference has failed to continue to meet the requirements for such preference, the Secretary may order the Director of Purchasing to: (a) reject the bid; or (b) assess a penalty against such Bidder in an amount not to exceed 5% of the bid amount and that such penalty will be paid to the contracting agency or deducted from any unpaid balance on the contract or purchase order.

By submission of this certificate, Bidder agrees to disclose any reasonably requested information to the Purchasing Division and authorizes the Department of Revenue to disclose to the Director of Purchasing appropriate information verifying that Bidder has paid the required business taxes, provided that such information does not contain the amounts of taxes paid nor any other information deemed by the Tax Commissioner to be confidential.

Under penalty of law for false swearing (West Virginia Code, §61-5-3), Bidder hereby certifies that this certificate is true and accurate in all respects; and that if a contract is issued to Bidder and if anything contained within this certificate changes during the term of the contract, Bidder will notify the Purchasing Division in writing immediately.

Bidder: _____ Signed: _____

Date: _____ Title: _____

*Check any combination of preference consideration(s) indicated above, which you are entitled to receive.