## SOLE SOURCE DETERMINATION

The Purchasing Division has been requested to approve a sole source purchase for the commodity or service described below. Pursuant to West Virginia Code 5A-3-10c, the Purchasing Division is attempting to determine whether the commodity or service is a sole source procurement. If you believe your company meets the required experience and qualification criteria stated below, please e-mail the Purchasing Division Buyer at <a href="Roberta.A.Wagner@wv.gov">Roberta.A.Wagner@wv.gov</a> with a copy to <a href="www.gov">w.michael.sheets@wv.gov</a> to express your interest in the project. Please forward any and all information that will support your company's compliance with required qualification and eligibility criteria along with any other pertinent information relative to this project to the Purchasing Division no later than 1:30 PM on 8/8/2011

Requisition Number: BPH12029

Department/Agency: Emergency Medical Services

Detailed Description of Project: The Office of Emergency Medical Services is in need of EZ-IO Intraosseous Infusion Systems to establish in ambulances as standardized equipment.

Proposed Sole Source Vendor: Vidacare Corporation

Specific Eligibility Criteria:

- •Power Driver must be a sealed, hand-held, lithium-battery powered medical drill.
- •The Power Drive must be capable of producing a minimum of 500 human insertions.
- •The Power Driver must be intended for the controlled insertion of an intraosseous needle set into human bone.
- •Needle tip must have four (4) asymmetrical bevels on the stylet, two (2) cutting tips on the catheter.
- •Needle tips must have black indicator lines located at 5mm increments from hub for accessing insertion depth.
- •The needle sets must be made of 304 stainless and catheters must be 15 gauge.

Specific Qualification Criteria:

•Vendor must be authorized to market, see and/or service this product in the State of West Virginia.