



State of West Virginia
 Department of Administration
 Purchasing Division
 2019 Washington Street East
 Post Office Box 50130
 Charleston, WV 25305-0130

Solicitation

NUMBER
CME14173

PAGE
1

ADDRESS CORRESPONDENCE TO ATTENTION OF
BOB KILPATRICK 304-558-0067

RFQ COPY

TYPE NAME/ADDRESS HERE

VENDOR

SHIP TO

HEALTH AND HUMAN RESOURCES
 BUREAU FOR PUBLIC HEALTH
 OFFICE CHIEF MEDICAL EXAMINER
 619 VIRGINIA STREET, WEST
 CHARLESTON, WV
 25302 304-558-4865

DATE PRINTED
05/19/2014

BID OPENING DATE: 05/28/2014

BID OPENING TIME 1:30PM

LINE	QUANTITY	UOP	CAT NO.	ITEM NUMBER	UNIT PRICE	AMOUNT
ADDENDUM NO. 1						
ADDENDUM ISSUED TO:						
1. PROVIDE REVISED SPECIFICATIONS, PER ATTACHMENT A						
2. TO CHANGE BID OPENING DATE AND TIME TO 5/28/2014 AT 1:30PM EST						
3. TO PROVIDE ADDENDUM ACKNOWLEDGEMENT. THIS DOCUMENT SHOULD BE SIGNED AND RETURNED WITH YOUR BID. FAILURE TO SIGN AND RETURN MAY RESULT IN THE DISQUALIFICATION OF YOUR BID.						
***** END OF ADDENDUM NO.1 *****						

SIGNATURE	TELEPHONE	DATE
TITLE	FEIN	ADDRESS CHANGES TO BE NOTED ABOVE

WHEN RESPONDING TO SOLICITATION, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'

SOLICITATION NUMBER: CME14173
Addendum Number: 1

The purpose of this addendum is to modify the solicitation identified as (“Solicitation”) to reflect the change(s) identified and described below.

Applicable Addendum Category:

- | Modify bid opening date and time
- | Modify specifications of product or service being sought
- | Attachment of vendor questions and responses
- | Attachment of pre-bid sign-in sheet
- | Correction of error
- | Other

Description of Modification to Solicitation:

Addendum #1 issued to modify specifications (per Attachment A) and to

Postpone bid Opening Date and Time to Wednesday, May 28, 2014 at 1:30pm EST.

NO OTHER CHANGES

Additional Documentation: Documentation related to this Addendum (if any) has been included herewith as Attachment A and is specifically incorporated herein by reference.

Terms and Conditions:

1. All provisions of the Solicitation and other addenda not modified herein shall remain in full force and effect.
2. Vendor should acknowledge receipt of all addenda issued for this Solicitation by completing an Addendum Acknowledgment, a copy of which is included herewith. Failure to acknowledge addenda may result in bid disqualification. The addendum acknowledgement should be submitted with the bid to expedite document processing.

ATTACHMENT A

ADDENDUM #1 – CME14173**ADDENDUM #1 TO REVISE RFQ SPECIFICATIONS AND EXTEND BID OPENING TO 1:30 PM, MAY 28, 2014.**

Revise language for Section 3.1.1.11 on the RFQ Specifications as follows:

Change From:

3.1.1.11 Quadrupole Time-of-Flight Mass Spectrometer must employ stacked ring, traveling wave ion cell after the quadrupole to help concentrate and focus ions before time of flight pusher.

Change To:

3.1.1.11 Quadrupole Time-of-Flight Mass Spectrometer must employ a technology method for transferring all ions rather than one ion at a time to help concentrate and focus ions that is equal to Waters' stacked ring "traveling wave" technology.

Revise language for Section 3.1.1.15 on the RFQ Specifications as follows:

Change From:

3.1.1.15 Quadrupole Time-of-Flight Mass Spectrometer must use a universal source architecture method to allow for source change and future upgrade.

Change To:

3.1.1.15 Quadrupole Time-of-Flight Mass Spectrometer must allow for source changes and future upgrades.

Revise language for Section 3.1.1.17 on the RFQ Specifications as follows:

Change From:

3.1.1.17 Quadrupole Time-of-Flight Mass Spectrometer must have an isolation valve to allow cleaning of cone without venting instrument.

Change To:

3.1.1.17 Quadrupole Time-of-Flight Mass Spectrometer must include a method to allow cleaning of cone without venting instrument.

Revise language for Section 3.1.1.30 on the RFQ Specifications as follows:

Change From:

3.1.1.30 Quadrupole Time-of-Flight Mass Spectrometer software must be capable of aligning elevated-energy mass spectrometry (MS^E) data according to liquid chromatography retention time and then automatically associating the precursor ion for each sample component to its fragment ion spectrum. It must be possible to automatically interrogate the processed elevated-energy mass spectrometry dataset.

Change To:

3.1.1.30 Quadrupole Time-of-Flight Mass Spectrometer software must be able to align parent and fragment data by retention time and allow user the ability to search one or both.

ADDENDUM ACKNOWLEDGEMENT FORM
SOLICITATION NO.: CME14173

Instructions: Please acknowledge receipt of all addenda issued with this solicitation by completing this addendum acknowledgment form. Check the box next to each addendum received and sign below. Failure to acknowledge addenda may result in bid disqualification.

Acknowledgment: I hereby acknowledge receipt of the following addenda and have made the necessary revisions to my proposal, plans and/or specification, etc.

Addendum Numbers Received:

(Check the box next to each addendum received)

- | | |
|---|--|
| <input type="checkbox"/> Addendum No. 1 | <input type="checkbox"/> Addendum No. 6 |
| <input type="checkbox"/> Addendum No. 2 | <input type="checkbox"/> Addendum No. 7 |
| <input type="checkbox"/> Addendum No. 3 | <input type="checkbox"/> Addendum No. 8 |
| <input type="checkbox"/> Addendum No. 4 | <input type="checkbox"/> Addendum No. 9 |
| <input type="checkbox"/> Addendum No. 5 | <input type="checkbox"/> Addendum No. 10 |

I understand that failure to confirm the receipt of addenda may be cause for rejection of this bid. I further understand that any verbal representation made or assumed to be made during any oral discussion held between Vendor's representatives and any state personnel is not binding. Only the information issued in writing and added to the specifications by an official addendum is binding.

Company

Authorized Signature

Date

NOTE: This addendum acknowledgement should be submitted with the bid to expedite document processing.
 Revised 6/8/2012