



State of West Virginia
 Department of Administration
 Purchasing Division
 2019 Washington Street East
 Post Office Box 50130
 Charleston, WV 25305-0130

Solicitation

NUMBER
WEH14006

PAGE
1

ADDRESS CORRESPONDENCE TO ATTENTION OF:
ROBERTA WAGNER 304-558-0067

RFQ COPY

TYPE NAME/ADDRESS HERE

VENDOR

SHIP TO

HEALTH AND HUMAN RESOURCES
 WELCH COMMUNITY HOSPITAL
 454 MCDOWELL STREET
 WELCH, WV
 24801 304-436-8710

DATE PRINTED
08/06/2013

BID OPENING DATE: 08/21/2013

BID OPENING TIME 1:30PM

LINE	QUANTITY	UOP	CAT NO.	ITEM NUMBER	UNIT PRICE	AMOUNT
				ADDENDUM NO. 1		
THIS ADDENDUM IS ISSUED:						
1. TO PROVIDE RESPONSES TO VENDORS' QUESTIONS REGARDING THE ABOVE SOLICITATION.						
2. TO PROVIDE ADDENDUM ACKNOWLEDGEMENT. THIS DOCUMENT SHOULD BE SIGNED AND RETURNED WITH YOUR BID. FAILURE TO SIGN AND RETURN MAY RESULT IN THE DISQUALIFICATION OF YOUR BID.						
***** END OF ADDENDUM NO. 1 *****						

SIGNATURE		TELEPHONE	DATE
TITLE	FEIN	ADDRESS CHANGES TO BE NOTED ABOVE	

WHEN RESPONDING TO SOLICITATION, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'



State of West Virginia
 Department of Administration
 Purchasing Division
 2019 Washington Street East
 Post Office Box 50130
 Charleston, WV 25305-0130

Solicitation

NUMBER
WEH14006

PAGE
2

ADDRESS CORRESPONDENCE TO ATTENTION OF
ROBERTA WAGNER 304-558-0067

RFQ COPY
 TYPE NAME/ADDRESS HERE

V
E
N
D
O
R

S
H
I
P
T
O

HEALTH AND HUMAN RESOURCES
 WELCH COMMUNITY HOSPITAL
 454 MCDOWELL STREET
 WELCH, WV
 24801 304-436-8710

DATE PRINTED
08/06/2013

BID OPENING DATE: 08/21/2013 BID OPENING TIME 1:30PM

LINE	QUANTITY	UOP	CAT NO	ITEM NUMBER	UNIT PRICE	AMOUNT
0001	8	EA		465-14 INTENSIVE CARE UNIT (ICU) BEDSIDE MONITORS.		
0002	10	EA		465-14 MEDICAL SURGICAL WEARABLE PATIENT MONITORS		
0003	1	EA		465-14 INFORMATION CENTER		
0004	18	EA		465-14 VITAL SIGNS MONITORS		
0005	1	EA		465-14 WARRANTY		

SIGNATURE	TELEPHONE	DATE
TITLE	FEIN	ADDRESS CHANGES TO BE NOTED ABOVE

WHEN RESPONDING TO SOLICITATION, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'



State of West Virginia
 Department of Administration
 Purchasing Division
 2019 Washington Street East
 Post Office Box 50130
 Charleston, WV 25305-0130

Solicitation

NUMBER
WEH14006

PAGE
3

ADDRESS CORRESPONDENCE TO ATTENTION OF:
ROBERTA WAGNER 304-558-0067

RFQ COPY
 TYPE NAME/ADDRESS HERE

VENDOR

SHIP TO

HEALTH AND HUMAN RESOURCES
 WELCH COMMUNITY HOSPITAL
 454 MCDOWELL STREET
 WELCH, WV
 24801 304-436-8710

DATE PRINTED
08/06/2013

BID OPENING DATE: 08/21/2013 BID OPENING TIME 1:30PM

LINE	QUANTITY	UOP	CAT NO	ITEM NUMBER	UNIT PRICE	AMOUNT
0006	1	EA		465-14		
				MANUAL/CDS		
0007	1	EA		465-14		
				INSTALLATION		
0008	1	EA		465-14		
				IN-SERVICE TRAINING		
***** THIS IS THE END OF RFQ WEH14006 ***** TOTAL:						

SIGNATURE		TELEPHONE	DATE
TITLE	FEIN	ADDRESS CHANGES TO BE NOTED ABOVE	

WHEN RESPONDING TO SOLICITATION, INSERT NAME AND ADDRESS IN SPACE ABOVE LABELED 'VENDOR'

SOLICITATION NUMBER: WEH14006

Addendum Number: 1

The purpose of this addendum is to modify the solicitation identified as (“Solicitation”) to reflect the change(s) identified and described below.

Applicable Addendum Category:

- Modify bid opening date and time
- Modify specifications of product or service being sought
- Attachment of vendor questions and responses
- Attachment of pre-bid sign-in sheet
- Correction of error
- Other

Description of Modification to Solicitation:

1. To provide responses to Vendors' questions.
2. To provide Addendum Acknowledgement.

Additional Documentation: Documentation related to this Addendum (if any) has been included herewith as Attachment A and is specifically incorporated herein by reference.

Terms and Conditions:

1. All provisions of the Solicitation and other addenda not modified herein shall remain in full force and effect.
2. Vendor should acknowledge receipt of all addenda issued for this Solicitation by completing an Addendum Acknowledgment, a copy of which is included herewith. Failure to acknowledge addenda may result in bid disqualification. The addendum acknowledgement should be submitted with the bid to expedite document processing.

ATTACHMENT A

WEH14006
Addendum 1

Q.1. There are (2) mentions of an Information Center in the packet.
One is 3.1.1.9 Which is under the "Bedside Monitor" Section.

The other is 3.1.3 Information Center, which seems to be its own individual system.

Please explain the difference between these (2) and if the first one is actually an additional center to the bedside monitors.

A.1. Yes, these are 2 different information centers, 3.1.1.9 is for the bedside monitors and 3.1.3 is for the wearable patient monitors. We have to be able to distinguish between the patients within our ICU that have the bedside monitors from the patients that are on the Medical/Surgical Unit that have the wearable patient monitors.

Q.2.a. For the Wearable Patient Monitors, I am assuming this is a monitor with a telemetry transmitter connection.
For example a Philips X2 Monitor with an attached Philips M2601B Telemetry Transmitter.
Would this be sufficient?

A.2.a. No, the X2 monitor coupled with a transmitter does not meet what is required. The patient wearable must have a viewing screen built into the wearable unit that is on the patient to allow the physician or nurse to view the cardiac rhythm, heart rate, and pulse ox from the screen located on the wearable monitor that is being worn by the patient. The impulse from the wearable is then transferred to the information center.

Q.2.b. Please explain the setup needed, whether it is a standalone wearable monitor, or a two piece system (monitor AND wearable transmitter).

A.2.b. This a standalone wearable monitor with a screen built in to the wearable monitor.

Q.3. For the Bedside Monitors, additional Modules will be needed to add the specific options needed such as CO2.
These will be add-ons to the monitor. Should we include this into (1) total price, or break them down individually?
There is only room for one total amount for the bedside monitors.

A.3. Vendor is to price each item as specified, if there are add-ons, the vendor should include additional pricing in the bid pricing for each.

Q.4. What is the total budget for this project, so we can assess the values/prices to cater to it.

A.4. This information cannot be provided.

Q.5. Can you please elaborate what specific make/model you need? We assume this is a wireless telemetry transmitter considering the term "Wearable."

However, "Surgical" and the need for "Blood Pressure" is making me wonder if this could mean an actual monitor that is portable with the patient maybe on a roll stand.

A.5. Intellivue MX 40 or equal.

Q.6. Will this Hospital accept certified refurbished equipment with warranties?
Or will it only accept ONLY brand new un-opened equipment?

A.6. New equipment only.

Q.7. For 3.1.3.4 , 3.1.3.5 , and 3.1.3.7 please clarify each of these and what they are. I understand 3.1.3.7 is the wearable patient monitor, but 3.1.3.4 and 3.1.3.5 are these the bedside monitors or a separate Telemetry system. There is no mention of a specific telemetry system needed.

A.7. 3.1.3.4 wearable patient monitor

3.1.3.5 vital signs monitor

3.1.3.7 wearable patient monitor

As described in the specifications, this is an information center that will be located at the nurse's station, not a bedside monitor. The patient wearable MX40 or equal, will transmit data to the information center. Please refer to Phillips IntelliVue M3150 information center for further details. Also, Phillips MX40 for the patient wearable.

ADDENDUM ACKNOWLEDGEMENT FORM
SOLICITATION NO.: WEH14006

Instructions: Please acknowledge receipt of all addenda issued with this solicitation by completing this addendum acknowledgment form. Check the box next to each addendum received and sign below. Failure to acknowledge addenda may result in bid disqualification.

Acknowledgment: I hereby acknowledge receipt of the following addenda and have made the necessary revisions to my proposal, plans and/or specification, etc.

Addendum Numbers Received:

(Check the box next to each addendum received)

- | | |
|---|--|
| <input type="checkbox"/> Addendum No. 1 | <input type="checkbox"/> Addendum No. 6 |
| <input type="checkbox"/> Addendum No. 2 | <input type="checkbox"/> Addendum No. 7 |
| <input type="checkbox"/> Addendum No. 3 | <input type="checkbox"/> Addendum No. 8 |
| <input type="checkbox"/> Addendum No. 4 | <input type="checkbox"/> Addendum No. 9 |
| <input type="checkbox"/> Addendum No. 5 | <input type="checkbox"/> Addendum No. 10 |

I understand that failure to confirm the receipt of addenda may be cause for rejection of this bid. I further understand that any verbal representation made or assumed to be made during any oral discussion held between Vendor's representatives and any state personnel is not binding. Only the information issued in writing and added to the specifications by an official addendum is binding.

Company

Authorized Signature

Date

NOTE: This addendum acknowledgement should be submitted with the bid to expedite document processing.
 Revised 6/8/2012